



Instructions

Clamp for curved and flat surfaces


1. Press quick release button while pulling down on cross bar to open clamp.




2. Turn v-plate to fit curved or flat surface.



3. Press quick release button while pushing up on cross bar to close clamp.





4. Tighten clamp into place by turning cross bar.



Clip

1. Slide A, flat edge up, over B.
2. Insert B into C.
3. Tighten A onto C by turning.



Watch our instructional video on freearmcare.com



**Do not put FreeArm in mouth.*

**Regularly clean FreeArm with disinfectant wipe.*


**Do not leave user unattended.*

**Not safe for use with MRI.*

**Bolus syringe, pump, feed bag and pole clamp are not included with FreeArm.*


Gravity Bolus Feeding

- Tighten clamp into place on curved or flat surface.
- Use clip to hold bolus syringe.
- Bend arm to desired height for feeding tolerance.



Pump Usage

- Tighten clamp into place on curved or flat surface.
- Attach pump to metal cylinder on arm using clamp included with your pump.
- Straighten arm. Bag should hang directly above bottom clamp.
- Attach clip. Hang feed bag from small hook on clip.



make meals and infusions easier